

extendSURE™ CONTROL

LOT SPECIFIC VALUE ASSIGNMENT CARD

LEVEL 1

Haemoglobin A1c	UNITS	MEAN	RANGE
NGSP ALIGNED IFCC ALIGNED	%	5.8	(4.8-6.8)
	mmol/mol	40	(34-46)
	g/dL	0.58	(0.48-0.68)
	mmol/L	0.36	(0.30-0.42)
Total Haemoglobin	g/dL	14.48	(11.87-17.09)
	mmol/L	8.98	(7.36-10.60)

MANUFACTURER'S INSTRUMENT/HAEMOGLOBIN A1c REAGENT ASSAY

BECKMAN COULTER AU CHEMISTRY SYSTEMS

For use with Beckman Coulter HbA1c Reagent Kit Part Number B00389 only

LOT 4247



Canterbury Scientific Ltd
71 Whiteligh Avenue
Addington, 8011
Christchurch
New Zealand

USE BY 2024-02-29



EC REP

EMERGO EUROPE
Prinsessegracht 20
2514 AP The Hague
The Netherlands



Version 09. Revision date 07/19

extendSURE™ CONTROL

LOT SPECIFIC VALUE ASSIGNMENT CARD

LEVEL 2

Haemoglobin A1c	UNITS	MEAN	RANGE
NGSP ALIGNED IFCC ALIGNED	%	10.2	(8.5-11.9)
	mmol/mol	87	(73-101)
	g/dL	1.23	(1.01-1.45)
	mmol/L	0.76	(0.62-0.90)
Total Haemoglobin	g/dL	13.85	(11.20-16.50)
	mmol/L	8.59	(6.95-10.23)

MANUFACTURER'S INSTRUMENT/HAEMOGLOBIN A1c REAGENT ASSAY

BECKMAN COULTER AU CHEMISTRY SYSTEMS

For use with Beckman Coulter HbA1c Reagent Kit Part Number B00389 only

LOT 4247



Canterbury Scientific Ltd
71 Whiteligh Avenue
Addington, 8011
Christchurch
New Zealand

USE BY 2024-02-29



EC REP

EMERGO EUROPE
Prinsessegracht 20
2514 AP The Hague
The Netherlands



Version 09. Revision date 07/19

Printed to either side of
225g card