

extendSURE™ CONTROL

LOT SPECIFIC VALUE ASSIGNMENT CARD

LEVEL 1

Haemoglobin A1c

NGSP ALIGNED

IFCC ALIGNED

UNITS

%

mmol/mol

MEAN

5.6

38

RANGE

(4.6-6.6)

(31-45)

MANUFACTURER'S INSTRUMENT/HAEMOGLOBIN A1c REAGENT ASSAY

BECKMAN COULTER **DxC AU** CHEMISTRY SYSTEMS

For use with Beckman Coulter HbA1c Advanced Reagent Kit Ref. Number B93009

LOT 4247

USE BY 2024-02-29

Ver 02. Revision date 07/19



Canterbury Scientific Ltd
71 Whiteleigh Avenue
Addington, 8011
Christchurch
New Zealand

EC REP

EMERGO EUROPE
Prinsessegracht 20
2514 AP The Hague
The Netherlands

extendSURE™ CONTROL

LOT SPECIFIC VALUE ASSIGNMENT CARD

LEVEL 2

Haemoglobin A1c

NGSP ALIGNED

IFCC ALIGNED

UNITS

%

mmol/mol

MEAN

10.0

86

RANGE

(8.3-11.7)

(72-100)

MANUFACTURER'S INSTRUMENT/HAEMOGLOBIN A1c REAGENT ASSAY

BECKMAN COULTER **DxC AU** CHEMISTRY SYSTEMS

For use with Beckman Coulter HbA1c Advanced Reagent Kit Ref. Number B93009

LOT 4247

USE BY 2024-02-29

Ver 02. Revision date 07/19



Canterbury Scientific Ltd
71 Whiteleigh Avenue
Addington, 8011
Christchurch
New Zealand

EC REP

EMERGO EUROPE
Prinsessegracht 20
2514 AP The Hague
The Netherlands

Printed to either side of
225g card