

extendSURE™

CONTROL

LOT SPECIFIC VALUE ASSIGNMENT CARD

LEVEL 1

Haemoglobin A1c	UNITS	MEAN	RANGE
NGSP ALIGNED IFCC ALIGNED	%	5.9	(4.9 - 6.9)
	mmol/mol	41	(34 - 48)
	g/dL	0.58	(0.48 - 0.68)
	mmol/L	0.36	(0.30 - 0.42)
Total Haemoglobin	g/dL	14.18	(11.80 - 16.56)
	mmol/L	8.79	(7.32 - 10.26)

MANUFACTURER'S INSTRUMENT/HAEMOGLOBIN A1c REAGENT ASSAY

BECKMAN COULTER DxC CHEMISTRY SYSTEMS
For use with Beckman Coulter HbA1c Reagent Kit Part Numbers A87905 and B36415

LOT 4204



EC REP

Canterbury Scientific Ltd
71 Whiteleigh Avenue
Addington, 8011
Christchurch
New Zealand



USE BY 2022-03-31

Version 10. Revision date 07/19



EMERGO EUROPE
Prinsessegracht 20
2514 AP The Hague
The Netherlands

extendSURE™

CONTROL

LOT SPECIFIC VALUE ASSIGNMENT CARD

LEVEL 2

Haemoglobin A1c	UNITS	MEAN	RANGE
NGSP ALIGNED IFCC ALIGNED	%	11.1	(9.3 - 12.9)
	mmol/mol	98	(82 - 114)
	g/dL	1.35	(1.13 - 1.57)
	mmol/L	0.84	(0.70 - 0.98)
Total Haemoglobin	g/dL	13.74	(11.32 - 16.16)
	mmol/L	8.52	(7.02 - 10.02)

MANUFACTURER'S INSTRUMENT/HAEMOGLOBIN A1c REAGENT ASSAY

BECKMAN COULTER DxC CHEMISTRY SYSTEMS
For use with Beckman Coulter HbA1c Reagent Kit Part Numbers A87905 and B36415

LOT 4204



EC REP

Canterbury Scientific Ltd
71 Whiteleigh Avenue
Addington, 8011
Christchurch
New Zealand



USE BY 2022-03-31

Version 10. Revision date 10/19



EMERGO EUROPE
Prinsessegracht 20
2514 AP The Hague
The Netherlands

Printed to either side of
225g card