

extendSURE™

CONTROL

LOT SPECIFIC VALUE ASSIGNMENT CARD

LEVEL 1

Haemoglobin A1c

NGSP ALIGNED

IFCC ALIGNED

UNITS

%

mmol/mol

g/dL

mmol/L

MEAN

6.0

42

0.58

0.36

14.06

8.72

RANGE

(5.0 - 7.0)

(36 - 48)

(0.48 - 0.68)

(0.30 - 0.42)

(11.45 - 16.67)

(7.10 - 10.34)

Total Haemoglobin

g/dL

mmol/L

MANUFACTURER'S INSTRUMENT/HAEMOGLOBIN A1c REAGENT ASSAY

BECKMAN COULTER AU CHEMISTRY SYSTEMS

For use with Beckman Coulter HbA1c Reagent Kit Part Number B00389 only

LOT

4324

CE 2797

EC REP



USE BY 2027-03-31



IVD

CH REP

Version 11. Revision date 11/23



Canterbury Scientific Ltd
71 Whiteleigh Avenue
Addington, 8011
Christchurch
New Zealand

EMERGO EUROPE
Westervortsedijk 60,
6827 AT Arnhem
The Netherlands

Beckman Coulter International S.A.
Rue Juste-Olivier 22,
1260 - Nyon 1 -
Switzerland

extendSURE™

CONTROL

LOT SPECIFIC VALUE ASSIGNMENT CARD

LEVEL 2

Haemoglobin A1c

NGSP ALIGNED

IFCC ALIGNED

UNITS

%

mmol/mol

g/dL

mmol/L

g/dL

mmol/L

MEAN

10.3

89

1.26

0.78

14.35

8.90

RANGE

(8.6 - 12.0)

(75 - 103)

(1.04 - 1.48)

(0.64 - 0.92)

(11.70 - 17.00)

(7.26 - 10.54)

Total Haemoglobin

MANUFACTURER'S INSTRUMENT/HAEMOGLOBIN A1c REAGENT ASSAY

BECKMAN COULTER AU CHEMISTRY SYSTEMS

For use with Beckman Coulter HbA1c Reagent Kit Part Number B00389 only

LOT

4324

CE 2797



USE BY 2027-03-31

Version 11. Revision date 11/23



IVD

EC REP

CH REP



Canterbury Scientific Ltd
71 Whiteleigh Avenue
Addington, 8011
Christchurch
New Zealand

EMERGO EUROPE
Westervootsedijk 60,
6827 AT Arnhem
The Netherlands

Beckman Coulter International S.A.
Rue Juste-Olivier 22,
1260 - Nyon 1 -
Switzerland