

extendSURE™

CONTROL

LOT SPECIFIC VALUE ASSIGNMENT CARD

LEVEL 1

Haemoglobin A1c

NGSP ALIGNED

IFCC ALIGNED

UNITS

%

mmol/mol

g/dL

mmol/L

MEAN

6.0

43

0.58

0.36

RANGE

(5.0 - 7.0)

(36 - 50)

(0.48 - 0.68)

(0.30 - 0.42)

Total Haemoglobin

g/dL

mmol/L

13.69

8.49

(11.31 - 16.07)

(7.02 - 9.96)

MANUFACTURER'S INSTRUMENT/HAEMOGLOBIN A1c REAGENT ASSAY

BECKMAN COULTER **DxC** CHEMISTRY SYSTEMS

For use with Beckman Coulter HbA1c Reagent Kit Part Numbers A87905 and B36415

LOT

4324



USE BY 2027-03-31

Version 12. Revision date 11/23



IVD



Canterbury Scientific Ltd
71 Whiteleigh Avenue
Addington, 8011
Christchurch
New Zealand

EC REP

EMERGO EUROPE
Westervoortsedijk 60,
6827 AT Arnhem
The Netherlands

extendSURE™

CONTROL

LOT SPECIFIC VALUE ASSIGNMENT CARD

LEVEL 2

Haemoglobin A1c

NGSP ALIGNED

IFCC ALIGNED

UNITS

%
mmol/mol
g/dL
mmol/L
g/dL
mmol/L

MEAN

10.8
95
1.30
0.81
13.73
8.52

RANGE

(9.0 - 12.6)
(79 - 111)
(1.08 - 1.52)
(0.67 - 0.95)
(11.31 - 16.15)
(7.02 - 10.02)

Total Haemoglobin

MANUFACTURER'S INSTRUMENT/HAEMOGLOBIN A1c REAGENT ASSAY

BECKMAN COULTER **DxC** CHEMISTRY SYSTEMS

For use with Beckman Coulter HbA1c Reagent Kit Part Numbers A87905 and B36415

LOT

4324



USE BY 2027-03-31

Version 12. Revision date 11/23



IVD



Canterbury Scientific Ltd
71 Whiteleigh Avenue
Addington, 8011
Christchurch
New Zealand

EC REP

EMERGO EUROPE
Westervoortsedijk 60,
6827 AT Arnhem
The Netherlands